



A Continuous-Discrete Extended Kalman Filter for State and Parameter Estimation in People with Type 1 Diabetes

Boiroux, Dimitri; Bátora, Vladimír; Hagdrup, Morten; Aradóttir, Tina Björk; Johannsen, Caroline; Tárník, Marian; Murgaš, Ján; Schmidt, Signe; Nørgaard, Kirsten; Poulsen, Niels Kjølstad

Total number of authors:
12

Publication date:
2015

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):
Boiroux, D., Bátora, V., Hagdrup, M., Aradóttir, T. B., Johannsen, C., Tárník, M., Murgaš, J., Schmidt, S., Nørgaard, K., Poulsen, N. K., Madsen, H., & Jørgensen, J. B. (2015). *A Continuous-Discrete Extended Kalman Filter for State and Parameter Estimation in People with Type 1 Diabetes*. Abstract from 19th Nordic Process Control Workshop, Trondheim, Norway.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

A Continuous-Discrete Extended Kalman Filter for State and Parameter Estimation in People with Type 1 Diabetes

Dimitri Boiroux^{1,2}, Vladimír Bátorá³, Morten Hagdrup¹, Tinna Björk Aradóttir¹, Caroline Johannsen¹, Marían Tárník³, Ján Murgaš³, Signe Schmidt^{2,4}, Kirsten Nørgaard⁴, Niels Kjølstad Poulsen¹, Henrik Madsen¹ and John Bagterp Jørgensen¹

¹ DTU Compute, Technical University of Denmark, Kgs. Lyngby, Denmark

² Danish Diabetes Academy, Odense, Denmark

³ Faculty of Electrical Engineering and Information Technology, Slovak University of Technology, Bratislava, Slovakia

⁴ Department of Endocrinology, Hvidovre Hospital, Denmark

Abstract

Closed-loop control of blood glucose, also called an artificial pancreas (AP), can free people with type 1 diabetes from the burden of managing their insulin therapy. In addition, it has the potential to provide tighter blood glucose control while reducing the risk of short- and long-term diabetes complications. Currently, linear model predictive control (MPC) algorithms are being tested for the AP and show promising results. However, the insulin-glucose dynamics are highly nonlinear, and would require advanced techniques for more accurate state and parameter estimation.

Here, we apply MPC to an AP for people with type 1 diabetes. We use a minimal identifiable physiological model for simulation. In particular, we estimate the model states and parameters using a continuous-discrete extended Kalman filter (EKF). The filter should be able handle uncertainties associated with meal intake and changes in insulin sensitivity. We describe the key aspects of the numerical implementation and provide quantitative insights into the factors limiting the achievement of acceptable AP performance.

Acknowledgements

Funded by the Danish Diabetes Academy supported by the Novo Nordisk Foundation.